



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

7m

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002264 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1.17.2007 Cal 1 0.102 AC Cal 2 0.102 AC


The instrument was placed into service on 1.17.2007.

Terry S. Gallegos

2.27.2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 27, day of February, 2007.



Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 8^{mm}
B4

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 29947 AGENCY Tucson Police DepartmentDATE 1.17.2007 TIME 1602INTOXILYZER SERIAL # 80-00 2264 LOCATION Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ✓✓
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed ✓
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 102 AC

Instrument is operating properly and accurately. YES ✓ NO

COMMENTS

SIGNATURE

Terry Gallegos

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
01/17/2007

16:05:50

2 of 8^m
Ph

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:06:51
Diagnostic Test	Pass	16:07:18
Air Blank	0.000	16:07:46
0.100 Cal Check	0.102	16:08:05
Air Blank	0.000	16:08:34
Subject Test	0.000	16:08:52
Air Blank	0.000	16:09:20
Five Minute Wait	ABT*	
Air Blank	0.000	16:09:55

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
01/17/2007

16:10:07

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:11:09
Diagnostic Test	Pass	16:11:36
Air Blank	0.000	16:12:04
0.100 Cal Check	0.101	16:12:22
Air Blank	IPS*	16:12:30
Air Blank	RFI**	16:12:38

*Improper Sample

**RFI Detect

Not a Successfully Completed Test Sequence

3 of 8^{mm}
BH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
01/17/2007

16:12:59

4 of 8
PH

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:13:50
Diagnostic Test	Pass	16:14:18
Air Blank	0.000	16:14:45
0.100 Cal Check	0.101	16:15:04
Air Blank	0.000	16:15:33
Subject Test	XXX*	16:16:02
Air Blank	0.000	16:16:35

*Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
01/17/2007

16:16:45

50f8^{ms}
TS4

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:17:37
Diagnostic Test	Pass	16:18:04
Air Blank	0.000	16:18:31
0.100 Cal Check	0.101	16:18:50
Air Blank	0.000	16:19:20
Subject Test	0.007	16:19:36
Air Blank	0.000	16:20:03
Five Minute Wait		
Air Blank	0.000	16:25:04
Subject Test	0.027*	16:28:12
Air Blank	0.000	16:28:46

*Deficient Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-002264

01/17/2007

16:29:22

6088^m
TS4

Standard Lot#: 629202I

Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	16:29:50
Cal Check	0.102	16:30:08
Air Blank	0.000	16:30:38


Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

7089^{ms}
DH

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 1.17.2007 TIME 1602
Intoxilyzer Serial # 80-002264 Location Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 102 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Terry Gallegos

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-002264

01/17/2007

16:02:52

80f8^{2nd}
B4

Standard Lot#: 629202I

Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	16:03:19
Cal Check	0.102	16:03:38
Air Blank	0.000	16:04:07


Operator's Signature